



CERTIFICATE OF CONFORMITY

Order code: **825-180**

RFX Gateway

Supplied with UK plug

Range: -50 to 300°C

Resolution: Dependant on indicator

Accuracy: $\pm 0.5^{\circ}\text{C}$ (from -10 to 100°C)

$\pm 1.5^{\circ}\text{C}$ from -50 to -20°C / 100 to 200°C

$\pm 2.5^{\circ}\text{C}$ from 200 to 300°C

Made in China

DECLARATION OF CONFORMITY

Conforms to RoHS & REACH Legislation

Also conforms to the following regulations; EC Reg 1935/2004, EU Reg 10/2011 & amendment 2020/1245, Directive 2011/65/EU & amendment 2015/863, EU Reg 2014/35/EU, EU Reg 2014/30/EU, EMC

We certify that the goods shown on this certificate are manufactured, tested, inspected and stored, to conform in all respects to the stated specifications and in accordance with the Quality System of ETI Limited, accredited to BS EN ISO 9001, as approved and audited by the British Standards Institution.

We also confirm that when used in an appropriate manner, the products are suitable for use within the Food Industry. Where appropriate and required, the product complies with EEC Directives with regard to European Legislation and CE Marking requirements. Including EC Regulation No 1935/2004 on Materials and articles intended to come into contact with food & Commission Regulation EU Reg 10/2011 & amendment 2020/1245 on plastic materials and articles intended to come into contact with food.

All ETI supplied instruments are suitable for use within HACCP plans where the quoted specification for critical control points match the performance of the instrument.

Signed. on behalf of Electronic Temperature Instruments Ltd

A handwritten signature in black ink, appearing to read 'A. Reid', is written over a faint circular stamp.

Andy Reid

Quality Manager

Valid as of date of purchase.

Riverside House
Easting Close
Worthing
West Sussex
BN14 8HQ
T +44 (0)1903 202151
E sales@etilttd.com
W etilttd.com