

To cancel "OUT" setting:

1. Under power off status, press down **MODE** and **TARE** at the same time for about 5 seconds → **Set-0**.
2. Press **TARE** for 4 times → **Set-4**.
3. Press **MODE** once → **10.00** the most right digit flashes. Please change the number 10.00 to 99.99.
--- Press **TARE** to change number 0~9.
--- Press **MODE** to shift position.
4. After changing the number to **99.99**, please press **MODE** and **TARE** at the same time → **GOOD** → **SEt-5**.
5. Switch the scale off and re-start. Re-calibrate it if weighing result is not accurate.

TELLIER S.A.S
au capital de 3.000.000
21, rue Guy Moquet
F- 95100 ARGENTEUIL
Tél. 01 34 11 38 38 - Fax 01 34 11 18 50
RCS Pontoise 492 599 386
TVA (VAT) FR 66 492 599 386